

**ZONMED****DECLARATION OF CONFORMITY****Holder of Certificate:** Shenzhen Zhongchuang 3D Technology Co., Ltd.**Address:** Room 204, Administration Building, Jia Tai Pharmaceutical, No.1 Industrial Zone, Tangwei Community, Fuhai Street, Bao'an District, Shenzhen City, China**Manufacturer:** Shenzhen Zhongchuang 3D Technology Co., Ltd.**Address:** Room 204, Administration Building, Jia Tai Pharmaceutical, No.1 Industrial Zone, Tangwei Community, Fuhai Street, Bao'an District, Shenzhen City, China**SRN** CN-MF-000048153**EU Authorized** Wiselink Medical (Ireland) Limited**Representative** Block 1 Unit 1&4, Northwood Court, Santry, Dublin, D09E348Ireland
Tel:+353 01 263 0888
Email: office@wiselinkie.com
SRN: IE-AR-000037361

Product information:	Product name	Model	Basic UDI-DI
	Tray Resin	PRS-R-10	697861181010MN
	Surgical Guide Resin	PRS-R-11	697861181012MS
	Smile Veneer Resin	PRS-R-07	697861181018N6
	IBT Resin	PRS-R-09	697861181013MU
	Try-in Resin	PRS-R-03	697861181007MZ
	Clear Aligner Resin	PRS-R-14	697861181016N2
	Temporary Veneer Resin	PRS-R-06	697861181019N8

Applicable Standard: EN ISO 13485:2016, EN ISO 14971:2021, EN ISO 15223-1:2021, EN 62366-1:2015, EN 1041:2008, EN ISO 1641:2009, ISO 10993-1:2020, EN ISO 10993-5:2009, ISO 10993-10:2021**Regulation:** (EU) 2017/745**Remark:** The declaration of conformity is based on an evaluation of the above mentioned product. Complies with the requirements of European Community MD regulation (EU) 2017/745. This declaration applies to all products manufactured in accordance with the attached manufacturing drawing which form part of this declaration.**Name:** Li Chongxiang**Title:** Deputy General Manager**Date:** 23/06/2025**Place of Issuance:** Shenzhen**Signature:**